



This nasal cannula, as supplied with the **EDS**, will need routine inspection and replacement. It should be inspected before each flight and replaced if cracked, plugged, or excessively soiled. For health reasons each person using the **EDS** should have their own cannula



00EDS-1086-00

SIEDS-0006-00d